



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:*

***Applus (Jiangsu) Testing Co., Ltd.***  
***No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China***

*and hereby declares that the Organization is accredited in accordance with  
the recognized International Standard:*

**ISO/IEC 17025:2017**

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

***Electrical Testing***  
***(As detailed in the supplement)***

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

*Initial Accreditation Date:*

*Issue Date:*

*Expiration Date:*

May 22, 2026

May 22, 2026

August 31, 2028

*Accreditation No.:*

*Certificate No.:*

136682

L26-441

Tracy Szerszen  
President

*The validity of this certificate is maintained through ongoing assessments based  
on a continuous accreditation cycle. The validity of this certificate should be  
confirmed through the PJLA website: [www.pjlab.com](http://www.pjlab.com)*

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China

Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Medical Electrical Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test, Immunity test	EN 60601-1 GB 9706.1 IEC 60601-1	Digital power meter Clamp on power logger AC power source AC power source Three-phase electric adjustable power supply Voltage regulator	F1, F2	F, O
Electrical	Environmental Test for Medical Electrical Equipment	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure	GB/T 14710	Drop tester, Walk-in high and low temperature alternating humidity and heat test chamber, Constant temperature and humidity test chamber, Vibration Test Systems Model, Shock table, Simulated motor transport test bench	F1, F2	F, O
Electrical	Degrees of Protection Provided by Enclosure (IP code)	Dust protection rating, Water protection rating	GB/T 4208 IEC 60529	Jointed test finger Power indicator Sphere Test rod $\Phi 2.5$ Test wire	F1, F2	F, O
Electrical	Medical Foot Switch	Operating force, Contact resistance, Mechanical strength, Mechanical life, Cable bending performance, Protection rating, Misoperation protection	YY/T 1057	Push pull gauge	F1, F2	F, O
Electrical	Application of Usability Engineering to Medical Devices	Formative usability testing	EN 62366-1 IEC 62366-1 YY/T 1474	Technical system	F1, F2	F, O
Electrical	Application of Usability Engineering to Medical Devices	Formative usability testing	EN 60601-1-6 IEC 60601-1-6 YY/T 9706.106	Technical system	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Alarm Systems in Medical Electrical Equipment and Medical Electrical Systems	Visual alarm brightness, Auditory alarm sound level, Alarm response time, Alarm priority, Alarm recognition distance, Alarm duration, Alarm fail-safe	EN 60601-1-8 IEC 60601-1-8 YY 9706.108	Spectral Irradiance Colorimeter User's Manual Stop watch, Oscilloscope, Oscilloscope, Sound level meter, Tapeline, Medical device alarm test system, Digital angle meter	F1, F2	F, O
Electrical	Environmentally Conscious Design	Environmentally conscious Testing	EN 60601-1-9 IEC 60601-1-9	Ground Resistance Tester, Multiple times pressure tester, AC withstand voltage tester	F1, F2	F, O
Electrical	Medical Electrical Equipment and Medical Electrical Systems Used in the Home Healthcare Environment	Protection rating, Drop height, Tilt stability, Power supply adaptability, Cable tensile strength, Temperature and humidity adaptability, Emission test Immunity test	EN 60601-1-11 IEC 60601-1-11 YY 9706.111	Digital power meter, Clamp on power logger, AC power source, Voltage regulator, Stop watch, DC power supply, Lithium battery high and low air pressure test chamber, Walk-in high and low temperature alternating humidity and heat test chamber, Constant temperature and humidity test chamber	F1, F2	F, O
Electrical	Medical Electrical Equipment and Medical Electrical Systems Intended for Use in the Emergency Medical Services Environment	Protection rating, Drop height, Tilt stability, Power supply adaptability, Cable tensile strength, Temperature and humidity adaptability, Alarm sound level, Earth impedance, Emission test Immunity test	EN 60601-1-12 IEC 60601-1-12 YY 9706.112	Digital power meter, Clamp on power logger, AC power source, Voltage regulator, Stop watch, DC power supply, Lithium battery high and low air pressure test chamber, Walk-in high and low temperature alternating humidity and heat test chamber, Constant temperature and humidity test chamber	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Endoscopic Equipment	Illuminance, Resolution, Pressure drop, security of attachment of a jet-injector to the rigid bronchoscope container pressure with an open-ended jet-injector bronchoscope, Minimum instrument channel width, Depth of field, Leakage current, Dielectric strength, Surface temperature, Mechanical strength, Waterproof rating, Alarm response time, Light source life, Operating force, Emission test, Immunity test	EN 60601-2-18 GB 9706.218 IEC 60601-2-18	Walk-in high and low temperature alternating humidity and heat test chamber, Constant temperature and humidity test chamber	F1, F2	F, O
Electrical	Medical Endoscopes and Endotherapy Devices	Pressure drop, security of attachment of a jet-injector to the rigid bronchoscope container pressure with an open-ended jet-injector bronchoscope, Minimum instrument channel width, Resolution, Depth of field, Surface temperature, Leakage current, Dielectric strength, Sealing performance, Protection rating	ISO 8600-1	Walk-in high and low temperature alternating humidity and heat test chamber, Constant temperature and humidity test chamber, Vertical Steam Pressure Sterilizer (Autoclave), Push-Pull Force Gauge, Constant temperature and humidity test chamber, RF Immunity Testing System	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Rigid Bronchoscopes	Pressure drop, security of attachment of a jet-injector to the rigid bronchoscope container pressure with an open-ended jet-injector bronchoscope, Minimum instrument channel width, Resolution, Depth of field, Distortion rate, Luminous flux, Illuminance, Color temperature, Color rendering index, Stray light coefficient	ISO 8600-2	Push-Pull Force Gauge, Constant temperature and humidity test chamber	F1, F2	F, O
Electrical	Rigid Bronchoscopes	Pressure drop, security of attachment of a jet-injector to the rigid bronchoscope container pressure with an open-ended jet-injector bronchoscope, Minimum instrument channel width, Resolution, Depth of field, Distortion rate, Luminous flux, Illuminance, Mechanical strength, Sealing performance, Corrosion resistance	YY/T 0069	Push-Pull Force Gauge, Constant temperature and humidity test chamber	F1, F2	F, O
Electrical	Medical Endoscopes and Endotherapy Devices	angle, Bending strength, Interface	ISO 8600-3	Automatic medical endoscope test system	F1, F2	F, O
Electrical	Medical Endoscopes and Endotherapy Devices	Diameter Size	ISO 8600-4	Digimatic caliper	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Medical Endoscopes and Endotherapy Devices	Resolution	ISO 8600-5	Automatic medical endoscope test system	F1, F2	F, O
Electrical	Medical Endoscopes and Endotherapy Devices	Water resistance	ISO 8600-7	Stop watch <input type="checkbox"/> Waterproofing test equipment	F1, F2	F, O
Electrical	Medical Endoscopes and Endotherapy Devices	Diameter, Length	ISO 8600-8	Automatic Imaging Instrument <input type="checkbox"/> Electrostatic Discharge Generator	F1, F2	F, O
Electrical	Rigid Endoscope	Distortion rate, Luminous flux, Illuminance, Color rendering index, Stray light coefficient, Field center deviation	YY 0068.1	Automatic medical endoscope test system	F1, F2	F, O
Electrical	Rigid Endoscope	diameter, Working length, Bending angle, Bending force, Mechanical strength, Interface durability, Sealing performance, Corrosion resistance, Impact resistance, Operating force, Connection firmness	YY/T 0068.2	Digimatic caliper, Tapeline, Digital angle meter	F1, F2	F, O
Electrical	Rigid Endoscope	Label durability, Packaging sealing performance, Transport vibration adaptability, Drop impact adaptability, Stacking load, Marking clarity	YY/T 0068.3	Vertical Steam Pressure Sterilizer (Autoclave)	F1, F2	F, O



## Certificate of Accreditation: Supplement

### Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Rigid Endoscope	Leakage current, Dielectric strength, Protective earth impedance, Surface temperature, Protection rating, Alarm response time, Mechanical safety	YY 0068.4	Walk-in high and low temperature alternating humidity and heat test chamber	F1, F2	F, O
Electrical	Rigid Resectoscope	Leakage current, Dielectric strength, Protective earth impedance, Surface temperature, Protection rating, Alarm response time, Mechanical safety	YY/T 0619	Automatic medical endoscope test system	F1, F2	F, O
Electrical	Fibre Optical Cables for Lighting	Size, Light Emission Angle	YY/T 0763	Tapeline	F1, F2	F, O
Electrical	General electronic and electrical products, Industrial scientific and medical products	Positioning and Coordination, Locking and Removal, Temperature, Corrosion preventive	YY/T 0842 ASTM D4169 YY/T 0681.15 ISO 16834 ISO 13355	Digimatic caliper, Tapeline, Drop tester, Walk-in high and low temperature alternating humidity and heat test chamber, Constant temperature and humidity test chamber, Vibration Test Systems Model, Shock table, Simulated motor transport test bench	F1, F2	F, O
Electrical	Endoscope surgical equipment - Shaver	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure, foot switch, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 0955	Vertical Steam Pressure Sterilizer (Autoclave), Push-Pull Force Gauge, Constant temperature and humidity test chamber	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China

Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Fiber Endoscopes	Distortion rate, Luminous flux, Illuminance, Color rendering index, Stray light coefficient, Field center deviation	YY/T 1028	Walk-in high and low temperature alternating humidity and heat test chamber, Constant temperature and humidity test chamber, Vertical Steam Pressure Sterilizer (Autoclave), Constant temperature and humidity test chamber	F1, F2	F, O
Electrical	Cold Light Sources	Radiation flux, Light uniformity, luminous flux, color rendering index, color temperature, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 1081	Total luminous flux measurement device for cold light source, Neutral electrode temperature rise tester	F1, F2	F, O
Electrical	Capsule Endoscopes	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Encapsulation and pressure resistance,	YY/T 1298	Walk-in high and low temperature alternating humidity and heat test chamber, Constant temperature and humidity test chamber, RF Immunity Testing System	F1, F2	F, O
Electrical	Video Endoscopes	Field of view, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 1587	Automatic medical endoscope test system	F1, F2	F, O
Electrical	Video Camera System	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 1603	Automatic medical endoscope test system	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Esophagoscope	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Temperature, Lux	YY/T 0070	Digimatic caliper, Tapeline	F1, F2	F, O
Electrical	Sigmoidoscopy and rectoscopy set	Length, Lux, Surface roughness test, Temperature, Corrosion resistance	YY/T 0071	Digimatic caliper, Tapeline	F1, F2	F, O
Electrical	Infusion Pumps and Controllers	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN 60601-2-24 GB 9706.224 IEC 60601-2-24	Leakage Current Tester, AC withstand voltage tester, Ground Resistance Tester	F1, F2	F, O
Electrical	Electrocardiographs	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN 60601-2-25 GB 9706.225 IEC 60601-2-25	Walk-in high and low temperature alternating humidity and heat test chamber, RF Immunity Testing System, Waterproofing test equipment	F1, F2	F, O
Electrical	Electroencephalographs	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN IEC 80601-2-26 GB 9706.226 IEC 80601-2-26	Neutral electrode temperature rise tester, Waterproofing test equipment, Multiple times pressure tester	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Electroencephalographic biofeedback equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test, Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure, Performance	YY/T 0903	Multiple times pressure tester, Constant temperature and humidity test chamber	F1, F2	F, O
Electrical	Electrocardiographic Monitoring Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN 60601-2-27 GB 9706.227 IEC 60601-2-27	Walk-in high and low temperature alternating humidity and heat test chamber, RF Immunity Testing System, Waterproofing test equipment	F1, F2	F, O
Electrical	Electrocardiographic Monitoring Equipment	Dielectric strength, contact resistance, connector holding force, noise level, defibrillation protection, tensile strength, flexural life, insertion/ removal cycles	YY 0828	AC withstand voltage tester, Stop watch, Temperature and humidity meter	F1, F2	F, O
Electrical	Ambulatory Electrocardiographic Systems	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN 60601-2-47 IEC 60601-2-47 YY 9706.247	Leakage Current Tester □ Ground Resistance Tester	F1, F2	F, O
Electrical	Multifunction Patient Monitoring Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN IEC 80601-2-49 IEC 80601-2-49 YY 9706.249	Leakage Current Tester □ Ground Resistance Tester	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China

Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Sphygmomanometers	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test, Immunity test, Performance	BS EN IEC 80601-2-30 EN IEC 80601-2-30 IEC 80601-2-30 YY 9706.230	Measurement Receiver, Artificial Mains Network (AMN) / LISN, Pulse Limiter, Voltage Probe, Frequency Divider, Switching Unit / Transfer Switch, Composite Antenna, Loop Antenna, Power Absorbing Device (RF Load), Fixture / Test Fixture, Harmonics, Flicker, and Voltage Interruption Test System, Harmonics, Flicker, and Voltage Interruption Power Source, Electrostatic Discharge (ESD) Generator, Electrostatic High-Voltage Meter, Three-Phase Burst Combiner, Capacitive Coupling Clamp, Three-Phase Surge Combination Generator, Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Power Frequency Magnetic Field Test System, Power Frequency Magnetic Field Test System, Coil / Helmholtz Coil, Signal Generator, Power Meter, Power Sensor / Power Probe, Power Sensor / Power Probe, Power Amplifier, Power Amplifier, Field Strength Probe, Horn Antenna, Log-Periodic Antenna, RF Immunity Test System, Electromagnetic Clamp, Electromagnetic Clamp Calibration Kit, Electromagnetic Clamp Calibration Kit, Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Current Injection Probe, Fixture / Test Fixture, CDN Calibration Kit (3-piece set), CDN Calibration Kit (3-piece set), CDN Calibration Kit (3-piece set), 6 dB Attenuator, CS (Conducted Susceptibility) Current Method Calibration Kit, CS (Conducted	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China

Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Sphygmomanometers	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test, Immunity test, Performance	BS EN IEC 80601-2-30 EN IEC 80601-2-30 IEC 80601-2-30 YY 9706.230	Susceptibility) Current Method Calibration Kit, Arbitrary Waveform Generator, Oscilloscope, Radiation Loop, Loop Magnetic Field Probe, Radiation Loop, Loop Magnetic Field Probe, Horn Antenna, Power Amplifier	F1, F2	F, O
Electrical	Sphygmomanometers	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test, Immunity test,	BS EN IEC 60601-2-34 IEC 60601-2-34 YY 9706.234 IEC 60601-2-33 EN IEC 60601-2-33 YY 9706.233 ANSI/AAMI 60601-2-33	Measurement Receiver, Artificial Mains Network (AMN) / LISN, Pulse Limiter, Voltage Probe, Frequency Divider, Switching Unit / Transfer Switch, Composite Antenna, Loop Antenna, Power Absorbing Device (RF Load), Fixture / Test Fixture, Harmonics, Flicker, and Voltage Interruption Test System, Harmonics, Flicker, and Voltage Interruption Power Source, Electrostatic Discharge (ESD) Generator, Electrostatic High-Voltage Meter, Three-Phase Burst Combiner, Capacitive Coupling Clamp, Three-Phase Surge Combination Generator, Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Power Frequency Magnetic Field Test System, Power Frequency Magnetic Field Test System, Coil / Helmholtz Coil, Signal Generator, Power Meter, Power Sensor / Power Probe, Power Sensor / Power Probe, Power Amplifier, Power Amplifier, Field Strength Probe, Horn Antenna, Log-Periodic Antenna, RF Immunity Test System, Electromagnetic Clamp, Electromagnetic Clamp Calibration Kit, Electromagnetic Clamp Calibration Kit, Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN),	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Sphygmomanometers	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test,	BS EN IEC 60601-2-34 IEC 60601-2-34 YY 9706.234 IEC 60601-2-33 EN IEC 60601-2-33 YY 9706.233 ANSI/AAMI 60601-2-33	Coupling/Decoupling Network (CDN), Current Injection Probe, Fixture / Test Fixture, CDN Calibration Kit (3-piece set), CDN Calibration Kit (3-piece set), CDN Calibration Kit (3-piece set), 6 dB Attenuator, CS (Conducted Susceptibility) Current Method Calibration Kit, CS (Conducted Susceptibility) Current Method Calibration Kit, Arbitrary Waveform Generator, Oscilloscope, Radiation Loop, Loop Magnetic Field Probe, Radiation Loop, Loop Magnetic Field Probe, Horn Antenna, Power Amplifier	F1, F2	F, O
Electrical	Clinical Thermometers	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test, Performance	EN ISO 80601-2-56 ISO 80601-2-56 YY 9706.256	Constant temperature and humidity test chamber, Digital temperature and atmosphere pressure meter, Standard Platinum Resistance Thermometer (SPRT)	F1, F2	F, O
Electrical	Clinical Electronic Thermometer	Temperature monitoring, repeatability testing, duration, current consumption, water resistance, tensile strength, corrosion resistance, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB/T 21416	Constant temperature and humidity test chamber, Digital temperature and atmosphere pressure meter, Standard Platinum Resistance Thermometer (SPRT)	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Clinical Infrared Thermometers	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure	GB/T 21417.1	Constant temperature and humidity test chamber, Digital temperature and atmosphere pressure meter, Standard Platinum Resistance Thermometer (SPRT)	F1, F2	F, O
Electrical	Pulse Oximeter Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test, Performance	EN ISO 80601-2-61 ISO 80601-2-61 YY 9706.261	High-Precision Fast Spectroradiometer, Neutral electrode temperature rise tester, Waterproofing test equipment	F1, F2	F, O
Electrical	Pulse Oximeter Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test, Performance	BS EN ISO 9919 ISO 9919 YY 0784	High-Precision Fast Spectroradiometer, Neutral electrode temperature rise tester, Waterproofing test equipment	F1, F2	F, O
Electrical	Operating Tables	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN IEC 60601-2-46 IEC 60601-2-46 YY 9706.246	Leakage Current Tester, AC withstand voltage tester, Multiple times pressure tester	F1, F2	F, O
Electrical	Electrically powered operating table	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure, force, Oscillating quantity	YY/T 1106	Digital power meter, Clamp on power logger, AC power source, AC power source, Three-phase electric adjustable power supply, Voltage regulator	F1, F2	F, O
Electrical	Medical Beds	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN 60601-2-52 IEC 60601-2-52 YY 9706.252	Medical bed impact test device	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Test of mechanical properties of furniture	Performance	GB/T 10357.6	Medical bed impact test device	F1, F2	F, O
Electrical	Rehabilitation training bed	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test, Force	GB/T 26340	Digimatic caliper, Tapeline	F1, F2	F, O
Electrical	General maternity beds	Length, Performance	YY/T 0045	Medical bed impact test device	F1, F2	F, O
Electrical	Rehabilitation training bed	Static load test ,Impact test, Lateral force test	GB/T 20403	Medical bed impact test device	F1, F2	F, O
Electrical	Surgical Luminaires and Luminaires for Diagnosis	Voltage ,Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN IEC 60601-2-41 IEC 60601-2-41 YY 9706.241	High-Precision Fast Spectroradiometer, Automatic medical endoscope test system	F1, F2	F, O
Electrical	Non-laser Light Source Equipment Intended for Therapeutic, Diagnostic, Monitoring and Cosmetic/Aesthetic Use	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN 60601-2-57 IEC 60601-2-57 YY 9706.257	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Photobiological Safety of Lamps and Lamp Systems	Light ,electromagnetic radiation	GB/T 20145 IEC 62471	Optical radiation system for medical electrical equipment	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Home Light Therapy Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Light, electromagnetic radiation Emission test, Immunity test	EN IEC 60601-2-83 GB 9706.283 IEC 60601-2-83	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Red Light Therapy Equipment	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure	YY/T 1496	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Infrared Ray Cauterization Therapy Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 0323	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Infrared devices for instant screening of human skin temperature	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test, Immunity test	GB/T 19146	Constant temperature and humidity test chamber, Digital temperature and atmosphere pressure meter, Standard Platinum Resistance Thermometer (SPRT)	F1, F2	F, O
Electrical	Ultraviolet Therapy Equipment	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure	YY/T 0901	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Medical LED equipment	Light, Irradiance Measurement	YY/T 1534	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Wearable devices	Light, Irradiance Measurement	GB/T 41265	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Surgical, Cosmetic, Therapeutic and Diagnostic Laser Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	EN IEC 60601-2-22 GB 9706.222 IEC 60601-2-22	Push pull gauge, Leakage Current Tester	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Laser Products	Laser, Radiation	EN 60825-1 GB/T 7247.1 IEC 60825-1	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Laser therapeutic equipment	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure, Laser, Radiation	YY/T 1751	Temperature and humidity meter, Multimeter	F1, F2	F, O
Electrical	Laser therapeutic equipment	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure, Laser, Radiation	YY 0845	Temperature and humidity meter, AC power source, Multimeter	F1, F2	F, O
Electrical	Laser therapeutic equipment	Laser, Radiation	YY 0846	Temperature and humidity meter, AC power source, Multimeter	F1, F2	F, O
Electrical	Laser therapeutic equipment	Laser, Radiation	YY 0307	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Laser therapeutic equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation,	GB/T 12257	Optical radiation system for medical electrical equipment	F1, F2	F, O
Electrical	Laser therapeutic equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Laser, Radiation	YY 0983	Optical radiation system for medical electrical equipment	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Dental Equipment	Voltage, Current, Power Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Laser, Radiation, Emission test Immunity test	EN IEC 80601-2-60 GB 9706.260 IEC 80601-2-60	AC withstand voltage tester, Leakage Current Tester	F1, F2	F, O
Electrical	Stationary Dental Units and Dental Patient Chairs	Voltage, Current, Power Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Laser, Radiation, Emission test Immunity test	ISO 7494-1 YY/T 1043.1	Test corner, Test corner, Data Acquisition System, Handheld temperature recorder	F1, F2	F, O
Electrical	Air, Water, Suction and Wastewater Systems	Pressure	ISO 7494-2 ISO 5467-2 YY/T 1043.2	Dental units pressure loss tester, Digital differential pressure meter	F1, F2	F, O
Electrical	Transportable Dental Units	Pressure	YY/T 1044	Dental units pressure loss tester	F1, F2	F, O
Electrical	Dental implant unit	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength	YY/T 1485	Torque Test Device for Dental Motors	F1, F2	F, O
Electrical	Root Canal Preparation Unit	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength	YY/T 1602	Digimatic caliper	F1, F2	F, O
Electrical	Dental Whitening Accelerator	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength	YY/T 1401	Optical radiation system for medical electrical equipment, Broadband Radiometer	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Powered Polymerization Activators	Radiation	ISO 10650 YY 0055	Test corner, Data Acquisition System, Handheld temperature recorder	F1, F2	F, O
Electrical	Handpieces and Motors	Emission test Immunity test	ISO 14457 YY 1045	Drop tester	F1, F2	F, O
Electrical	Electric dental engine	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Rotation rate,	YY/T 1147	Stop watch, Voltage regulator, AC withstand voltage tester, DC power supply, Torque Test Device for Dental Motors, Tachometers	F1, F2	F, O
Electrical	Operating Lights	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, illuminance, illumination uniformity	ISO 9680 YY/T 1120	Leakage Current Tester, Ground Resistance Tester, Neutral electrode temperature rise tester	F1, F2	F, O
Electrical	Hot Melt Gutta Percha Filling Machine	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Emission test, Immunity test	YY/T 1692	Walk-in high and low temperature alternating humidity and heat test chamber, Electrostatic Discharge Generator, Lithium battery high and low air pressure test chamber	F1, F2	F, O
Electrical	Handpiece connectors	Size Measurement, Bonding force	YY/T 1012	Automatic Imaging Instrument	F1, F2	F, O
Electrical	Apex Locator	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 1684	AC withstand voltage tester, Leakage Current Tester, Digimatic caliper	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Medical Robots for Rehabilitation, Assessment, Compensation, or Alleviation	Electro Magnetic Compatibility Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	EN IEC 80601-2-78 IEC 80601-2-78 YY 9706.278	AC withstand voltage tester, Multiple times pressure tester, Digital angle meter, RF Immunity Testing System	F1, F2	F, O
Electrical	Motion rehabilitation training robot	Electro Magnetic Compatibility Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Stabilization, Noise, Fatigue strength, Emission test, Immunity test	GB/T 37704	Digimatic caliper, Tapeline, Digital angle meter, Automatic Imaging Instrument	F1, F2	F, O
Electrical	Walk training platform with suspension	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Stabilization, Noise	YY/T 0900	Tapeline	F1, F2	F, O
Electrical	Rehabilitation training instrument	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Stabilization, Static strength, Noise	GB 24436	Tapeline, Digimatic caliper	F1, F2	F, O
Electrical	Rehabilitation training instrument	Static strength, Stabilization	GB/T 26346	Tapeline, Digimatic caliper, Digital angle meter	F1, F2	F, O
Electrical	Rehabilitation training instrument	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Stabilization, strength,	GB/T 28919	Digimatic caliper, Tapeline, Digital angle meter	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Balance evaluation and exercise systems	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Stabilization, Noise, Emission test, Immunity test	YY/T 1410	Leakage Current Tester, AC withstand voltage tester, Multiple times pressure tester, RF Immunity Testing System	F1, F2	F, O
Electrical	Elbow, knee joints passive motion equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Time, fatigue tests, Noise	YY/T 0997	AC withstand voltage tester, Leakage Current Tester	F1, F2	F, O
Electrical	Table, chair and accessories used for electric cervical and lumbar traction therapy	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Pulley Stability, Emission test Immunity test	YY/T 1491	Medical bed impact test device	F1, F2	F, O
Electrical	Electrical upper/lower limbs round exercise therapy equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Rotation rate, Emission test Immunity test	YY/T 1626	Tachometers □ Push-Pull Force Gauge □ Tapeline	F1, F2	F, O
Electrical	High Frequency Surgical Equipment and High Frequency Surgical Accessories	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Monitoring NE, Emission test Immunity test	EN IEC 60601-2-2 GB 9706.202 IEC 60601-2-2	Digital power meter, AC power source, AC power source, Voltage regulator	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Plasma surgical equipment	Working Frequency, Power, Noise, Foot Switch □ Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure Emission test, Immunity test	YY/T 1409	Oscilloscope	F1, F2	F, O
Electrical	Electrical surgical equipment for osseous tissue	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Rotation rate, Foot switch	YY/T 0752	Tachometers, Temperature and humidity meter, Sound level meter, Tapeline	F1, F2	F, O
Electrical	Extracorporeally Induced Lithotripsy	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test, Immunity test	EN 60601-2-36 GB 9706.236 IEC 60601-2-36	Leakage Current Tester, AC withstand voltage tester, Ground Resistance Tester	F1, F2	F, O
Electrical	Robot-Assisted Surgical Systems	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test, Immunity test	EN IEC 80601-2-77 IEC 80601-2-77 YY 9706.277	Walk-in high and low temperature alternating humidity and heat test chamber	F1, F2	F, O
Electrical	Robot-Assisted Surgical Systems	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 1712	Walk-in high and low temperature alternating humidity and heat test chamber	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Radio-frequency Ablation Equipment for Ear, Nose and Throat	Foot switch, Working frequency, Time, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	YY 0897	Oscilloscope, Electrosurgical Unit Analyzer	F1, F2	F, O
Electrical	Radio Frequency Ablation Catheter	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY 0778	Oscilloscope, Electrosurgical Unit Analyzer	F1, F2	F, O
Electrical	Radio Frequency Ablation Equipment	Output wave form, Output power, Time, Foot switch, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY 0650	Oscilloscope, Electrosurgical Unit Analyzer	F1, F2	F, O
Electrical	Radio frequency ablation instrument for cardiac application	Foot switch, Working frequency, Time, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Emission test Immunity test	YY/T 0860	Oscilloscope, Electrosurgical Unit Analyzer	F1, F2	F, O
Electrical	Radio Frequency Ablation Instrument for Liver	Rotation rate, Foot switch, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Emission test Immunity test	YY/T 0776	Oscilloscope, Electrosurgical Unit Analyzer	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Ultrasonic Physiotherapy Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Emission test Immunity test	IEC 60601-2-5 EN 60601-2-5 GB 9706.205	Digital power meter, Clamp on power logger, AC power source, AC power source, Three-phase electric adjustable power supply	F1, F2	F, O
Electrical	Ultrasonic Medical Diagnostic and Monitoring Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation <input type="checkbox"/> Emission test, Immunity test	IEC 60601-2-37 EN IEC 60601-2-37 GB 9706.237	Digital power meter, Clamp on power logger, AC power source, AC power source, Three-phase electric adjustable power supply	F1, F2	F, O
Electrical	Electroacupuncture therapy device	Direct component, Time of therapy, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation <input type="checkbox"/> Emission test, Immunity test	YY 0780	Oscilloscope	F1, F2	F, O
Electrical	Electric heating moxibustion equipment	Temperature, work noise, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 1490	Data Acquisition System	F1, F2	F, O
Electrical	Steam therapy apparatus	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure, Emission test Immunity test	YY/T 1306	Stop watch, Precision constant temperature water bath, Multi-functional precision thermometer, Data Acquisition System, Handheld temperature recorder	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Short-wave Therapy Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Emission test Immunity test	EN 60601-2-3 GB 9706.203 IEC 60601-2-3	Leakage Current Tester □ Ground Resistance Tester	F1, F2	F, O
Electrical	Microwave Therapy Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Emission test Immunity test	EN 60601-2-6 GB 9706.206 IEC 60601-2-6	Leakage Current Tester, AC withstand voltage tester, Ground Resistance Tester	F1, F2	F, O
Electrical	Nerve and Muscle Stimulators	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Emission test Immunity test	EN 60601-2-10 IEC 60601-2-10 YY 9706.210 YY/T 0696	Digital power meter, Clamp on power logger, AC power source, AC power source, Three-phase electric adjustable power supply, Voltage regulator	F1, F2	F, O
Electrical	Electromyographs and Evoked Response Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Emission test Immunity test	EN 60601-2-40 IEC 60601-2-40 YY 9706.240	Digital power meter, Clamp on power logger, AC power source, AC power source, Three-phase electric adjustable power supply	F1, F2	F, O
Electrical	Oxygen Concentrator Equipment	Oxygen concentration accuracy, Pressure, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, performance, Emission test Immunity test	EN ISO 80601-2-69 ISO 80601-2-69 YY 9706.269	AC withstand voltage tester, Leakage Current Tester	F1, F2	F, O
Electrical	In vitro diagnostic medical devices	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 1454	AC withstand voltage tester, Leakage Current Tester	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Specified Electromagnetic Wave Therapy Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY/T 0061	Neutral electrode temperature rise tester	F1, F2	F, O
Electrical	Heat Radiation Therapy Equipment	Heat radiation, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	YY 0306	Precision constant temperature water bath	F1, F2	F, O
Electrical	Radio frequency hyperthermia equipment	Temperature, Humidity, Vibration, Shock, Stacking load, Transportation performance, Power supply adaptability, Atmospheric pressure, Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Working frequency, Emission test Immunity test	YY 0777	Power meter, Power sensor	F1, F2	F, O
Electrical	High Frequency Electrocautery Therapy Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation Emission test, Immunity test	YY/T 0322	Oscilloscope	F1, F2	F, O
Electrical	Sleep apnoea breathing therapy equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, Respiratory airways, Emission test, Immunity test	YY 9706.270 ISO 80601-2-70 EN ISO 80601-2-70	Leakage Current Tester, AC withstand voltage tester, Multiple times pressure tester, RF Immunity Testing System	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Photodynamic therapy and photodynamic diagnostic equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation, wavelength, light output Emission test, Immunity test	GB 9706.275 IEC 60601-2-75 EN IEC 60601-2-75	Leakage Current Tester, AC withstand voltage tester, Multiple times pressure tester, RF Immunity Testing System	F1, F2	F, O
Electrical	Medical Electrical Equipment	Emission test Immunity test	EN 60601-1-2 IEC 60601-1-2 YY 9706.102	Measurement Receiver, Artificial Mains Network (AMN) / LISN, Pulse Limiter, Voltage Probe, Frequency Divider, Switching Unit / Transfer Switch	F1, F2	F, O
Electrical	Medical Electrical Equipment	Immunity test	IEC TS 60601-4-2	Electrostatic Discharge (ESD) Generator, Electrostatic High-Voltage Meter	F1, F2	F, O
Electrical	General electronic and electrical products	Electrical fast transient/bursts immunity test	EN 61000-4-4 GB/T 17626.4 IEC 61000-4-4	Three-Phase Burst Combiner, Capacitive Coupling Clamp	F1, F2	F, O
Electrical	General electronic and electrical products	Power frequency magnetic field immunity test	EN 61000-4-8 GB/T 17626.8 IEC 61000-4-8	Power Frequency Magnetic Field Test System, Power Frequency Magnetic Field Test System, Coil / Helmholtz Coil	F1, F2	F, O
Electrical	General electronic and electrical products	ESD immunity test	EN IEC 61000-4-2 GB/T 17626.2 IEC 61000-4-2	Electrostatic Discharge (ESD) Generator, Electrostatic High-Voltage Meter	F1, F2	F, O
Electrical	General electronic and electrical products	surge immunity test	EN 61000-4-5 GB/T 17626.5 IEC 61000-4-5	Three-Phase Surge Combination Generator, Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN)	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China

Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	General electronic and electrical products	Radiated field in close proximity immunity test	GB/T 17626.39 EN 61000-4-39 IEC 61000-4-39	Switching Unit / Transfer Switch, Signal Generator, Power Meter, Power Sensor / Power Probe, Power Sensor / Power Probe, Power Amplifier, Power Amplifier, Field Strength Probe, Arbitrary Waveform Generator, Oscilloscope, Radiation Loop, Loop Magnetic Field Probe, Radiation Loop, Loop Magnetic Field Probe, Horn Antenna, Power Amplifier	F1, F2	F, O
Electrical	General electronic and electrical products	Conducted immunity test	EN IEC 61000-4-6 GB/T 17626.6 IEC 61000-4-6	RF Immunity Test System, Electromagnetic Clamp, Electromagnetic Clamp Calibration Kit, Electromagnetic Clamp Calibration Kit, Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Coupling/Decoupling Network (CDN), Current Injection Probe, Fixture / Test Fixture, CDN Calibration Kit (3-piece set), CDN Calibration Kit (3-piece set), CDN Calibration Kit (3-piece set), 6 dB Attenuator, CS (Conducted Susceptibility) Current Method Calibration Kit, CS (Conducted Susceptibility) Current Method Calibration Kit	F1, F2	F, O
Electrical	General electronic and electrical products	Radiated immunity test	EN IEC 61000-4-3 GB/T 17626.3 IEC 61000-4-3	Switching Unit / Transfer Switch, Signal Generator, Power Meter, Power Sensor / Power Probe, Power Sensor / Power Probe, Power Amplifier, Power Amplifier, Field Strength Probe, Horn Antenna, Log-Periodic Antenna	F1, F2	F, O
Electrical	General electronic and electrical products	Harmonic current emission test	EN IEC 61000-3-2 GB 17625.1 IEC 61000-3-2	Harmonics, Flicker, and Voltage Interruption Test System, Harmonics, Flicker, and Voltage Interruption Power Source	F1, F2	F, O



## Certificate of Accreditation: Supplement

### Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	General electronic and electrical products	Voltage fluctuations and flicker test	EN 61000-3-3 GB/T 17625.2 IEC 61000-3-3	Harmonics, Flicker, and Voltage Interruption Test System, Harmonics, Flicker, and Voltage Interruption Power Source	F1, F2	F, O
Electrical	General electronic and electrical products	voltage dips short interruptions and voltage variation immunity test	EN IEC 61000-4-11 GB/T 17626.11 IEC 61000-4-11	Harmonics, Flicker, and Voltage Interruption Test System, Harmonics, Flicker, and Voltage Interruption Power Source	F1, F2	F, O
Electrical	Industrial, scientific and medical equipment	Emission test	CISPR 11 EN IEC 55011 GB 4824	Measurement Receiver, Artificial Mains Network (AMN) / LISN, Pulse Limiter, Voltage Probe, Frequency Divider, Switching Unit / Transfer Switch	F1, F2	F, O
Electrical	Industrial, scientific and medical equipment	Immunity test	GB/T 38326	Electrostatic Discharge (ESD) Generator, Static Gun Target Set, Static Gun Target Set, Electrostatic High-Voltage Meter	F1, F2	F, O
Electrical	Industrial, scientific and medical equipment	Emission test	GB/T 38336	Harmonics, Flicker, and Voltage Interruption Test System, Harmonics, Flicker, and Voltage Interruption Power Source	F1, F2	F, O
Electrical	Wheelchairs	Emission test Immunity test	GB/T 18029.21 ISO 7176-21	Measurement Receiver, Artificial Mains Network (AMN) / LISN, Pulse Limiter, Voltage Probe, Frequency Divider, Switching Unit / Transfer Switch	F1, F2	F, O
Electrical	Industrial, scientific and medical equipment	Emission test	CISPR 15	Measurement Receiver, Artificial Mains Network (AMN) / LISN, Pulse Limiter, Voltage Probe, Frequency Divider, Switching Unit / Transfer Switch	F1, F2	F, O
Electrical	Industrial, scientific and medical equipment	Emission test	GB/T 17743	Measurement Receiver, Artificial Mains Network (AMN) / LISN, Pulse Limiter, Voltage Probe, Frequency Divider, Switching Unit / Transfer Switch	F1, F2	F, O



## Certificate of Accreditation: Supplement

### Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Electrical Equipment for Measurement, Control, and Laboratory Use	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB 4793	Digital power meter, Clamp on power logger, Oscilloscope, AC withstand voltage tester, Voltage regulator, Digital angle meter, Multimeter, Data Acquisition System, Stop watch	F1, F2	F, O
Electrical	Electrical Equipment for Measurement, Control, and Laboratory Use	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	EN 61010-1 GB 4793.1 GB/T 42125.1 IEC 61010-1	Digital power meter, Clamp on power logger, Oscilloscope, AC withstand voltage tester, Voltage regulator, Digital angle meter, Multimeter, Data Acquisition System, Stop watch	F1, F2	F, O
Electrical	Laboratory Equipment for the Heating of Materials	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB/T 42125.2 EN IEC 61010-2-010 IEC 61010-2-010 GB 4793.6	Digital power meter, Clamp on power logger, Oscilloscope, AC withstand voltage tester, Voltage regulator, Digital angle meter, Multimeter, Data Acquisition System, Stop watch	F1, F2	F, O
Electrical	Laboratory Centrifuges	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	EN 61010-2-020 IEC 61010-2-020 GB 4793.7 GB/T 42125.5	Digital power meter, Clamp on power logger, Oscilloscope, AC withstand voltage tester, Voltage regulator, Digital angle meter, Multimeter, Data Acquisition System, Stop watch	F1, F2	F, O
Electrical	Sterilizers and Washer-disinfectors Used to Treat Medical Materials	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	EN IEC 61010-2-040 IEC 61010-2-040 GB 4793.4	Digital power meter, Clamp on power logger, Oscilloscope, AC withstand voltage tester, Voltage regulator, Digital angle meter, Multimeter, Data Acquisition System, Stop watch	F1, F2	F, O
Electrical	Laboratory Equipment for Mixing and Stirring	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	EN IEC 61010-2-051 IEC 61010-2-051 GB 4793.3	Digital power meter, Clamp on power logger, Oscilloscope, AC withstand voltage tester, Voltage regulator, Digital angle meter, Multimeter, Data Acquisition System, Stop watch	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Automatic and Semi-automatic Laboratory Equipment for Analysis and Other Purposes	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	EN IEC 61010-2-081 IEC 61010-2-081 GB 4793.9 GB/T 42125.14	Digital power meter, Clamp on power logger, Oscilloscope, AC withstand voltage tester, Voltage regulator, Digital angle meter, Multimeter, Data Acquisition System, Stop watch	F1, F2	F, O
Electrical	In Vitro Diagnostic (IVD) Medical Equipment	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	EN IEC 61010-2-101 IEC 61010-2-101 YY 0648	Digital power meter, Clamp on power logger, Oscilloscope, AC withstand voltage tester, Voltage regulator, Digital angle meter, Multimeter, Data Acquisition System, Stop watch	F1, F2	F, O
Electrical	Electrical Equipment for Measurement, Control and Laboratory	Emission test Immunity test	EN IEC 61326-1 GB/T 18268.1 IEC 61326-1	Measurement Receiver, Artificial Mains Network (AMN) / LISN	F1, F2	F, O
Electrical	Electrical Equipment for Measurement, Control and Laboratory	Emission test Immunity test	EN IEC 61326-2-6 GB/T 18268.26 IEC 61326-2-6	Measurement Receiver, Artificial Mains Network (AMN) / LISN	F1, F2	F, O
Electrical	Audio, video, information technology and communication technology equipment	Voltage, Current, Power, Insulation, Security, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB 4943.1 IEC 81001-5-1 IEC 81001-2-2 IEC 62368-1 EN IEC 62368-1 BS EN IEC 62368-1 AS/NZS 62368.1 UL 62368-1	Leakage Current Tester Three-phase pulse group combiner Winding temperature rise tester Neutral electrode temperature rise tester UV-Vis-NIR Spectral Analysis System Constant temperature and humidity test chamber Horizontal and vertical combustion tester Ball pressure tester, Drop tester Spring hammer, Vibration Test Systems Model Laser Power Meter Arbitrary Signal Generator	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Safety of household and similar electrical appliances	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB 4706.1 IEC 60335-1 EN 60335-1 BS EN 60335-1 BS EN IEC 60335-1 AS/NZS 60335.1 UL 60335-1 GB/T 4706.1 EN IEC 60335-1	Multiple times pressure tester Ground Resistance Tester Ball pressure tester Spring hammer Waterproofing test equipment Horizontal and vertical combustion tester	F1, F2	F, O
Electrical	Grills, toasters and similar portable cooking appliances	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB 4706.14 IEC 60335-2-9 EN 60335-2-9 EN IEC 60335-2-9 AS/NZS 60335.2.9 GB/T 4706.14	Dust test chamber Spring hammer Waterproofing test equipment Horizontal and vertical combustion tester Multiple times pressure tester Ground Resistance Tester Ball pressure tester	F1, F2	F, O
Electrical	Liquid heaters	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB 4706.19 IEC 60335-2-15 EN 60335-2-15 BS EN 60335-2-15 AS/NZS 60335.2.15 GB/T 4706.19	Multiple times pressure tester Ground Resistance Tester Ball pressure tester Spring hammer Waterproofing test equipment Horizontal and vertical combustion tester Constant temperature and humidity test chamber Horizontal and vertical combustion tester	F1, F2	F, O



# Certificate of Accreditation: Supplement

## Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China  
 Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Electrical	Fans	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB 4706.27 IEC 60335-2-80 EN 60335-2-80 BS EN 60335-2-80 BS EN IEC 60335-2-80 AS/NZS 60335.2.80 GB/T 4706.27 EN IEC 60335-2-80	Multiple times pressure tester Ground Resistance Tester Constant temperature and humidity test chamber Ball pressure tester Spring hammer Horizontal and vertical combustion tester	F1, F2	F, O
Electrical	Battery chargers	Voltage, Current, Power, Insulation, Earth impedance, Leakage current, Temperature, Mechanical load, Dielectric strength, Radiation	GB 4706.18 IEC 60335-2-29 EN 60335-2-29 AS/NZS: 60335.2.29 GB/T 4706.18	Multiple times pressure tester Ground Resistance Tester Constant temperature and humidity test chamber Ball pressure tester Spring hammer Waterproofing test equipment Oscilloscope	F1, F2	F, O
Electrical	Medical Electrical Equipment	Methods for product accelerated testing	GB/T 34986 IEC 62506	Constant temperature and humidity test chamber Walk-in high and low temperature alternating humidity and heat test chamber Vibration Test Systems Model	F1, F2	F, O
Electrical	Medical instruments of stainless steel	Corrosion resistance	YY/T 0149 ISO 13402	Vertical Steam Pressure Sterilizer (Autoclave)	F1, F2	F, O



## Certificate of Accreditation: Supplement

### Applus (Jiangsu) Testing Co., Ltd.

No.465, Lucai North Road, Huaxin Town, QingPu District, Shanghai, China

Contact Name: Sherry Zhu Phone: +86 151-6116-5605

*Accreditation is granted to the facility to perform the following conformity assessment activities:*

1. Location of activity:

**Location**

F

Conformity assessment activity is performed at the CABs fixed facility

O

Conformity assessment activity is performed onsite at the CABs customer location

**Location**

2. Flex Code:

F0- Fixed scope item. No deviations allowed to the line item as identified, except for updating to the most recent version of an accredited standard method after verification.

F1- Laboratory has the capability to test a new item, material, matrix, or product similar in composition to item, material, matrix, or product identified on the scope

F2- Laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope

F3- Laboratory has the capability to introduce a parameter/component/analyte to an accredited test method identified on the scope

F4- Laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope

F5- Laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using same technology or technique) identified on the scope

